

## EPA Método 608 Pesticidas (Confirmação)

Temperature: 150° (6 min. hold) (5°/min.) - 290°C

CGS 1027.14

Injector: 250°C  
Detector: 325°C, ECD  
Carrier Gas: 22 cm/sec., Helium

- 1-  $\alpha$ -BHC
- 2-  $\gamma$ -BHC
- 3-  $\beta$ -BHC
- 4- Heptachlor
- 5- BHC
- 6- Aldrin
- 7- Heptachlor Epoxide
- 8- Endosulfan I
- 9- p,p'-DDE
- 10- Dieldrin
- 11- Endrin
- 12- Endosulfan II
- 13- p,p'-DDD
- 14- Endrin Aldehyde
- 15- p,p'-DDT
- 16- Endosulfan Sulfate

