

## EPA Método 625 Base Neutra

Temperature: 40° (5 min. hold) (6°/min.) - 280°C

Injector: 230°C  
 Detector: MSD  
 Carrier Gas: 30 cm/sec., Helium

1- n-nitrosodimethylamine  
 2- bis(2-chloroethyl) ether  
 3- 1,3-dichlorobenzene  
 4- 1,4-dichlorobenzene  
 5- 1,2-dichlorobenzene  
 6- bis(2-chloroisopropyl) ether  
 7- hexachloroethane  
 8- n-nitrosodipropylamine  
 9- nitrobenzene  
 10- isophorone  
 11- bis(2-chloroethoxy) methane  
 12- 1,2,4-trichlorobenzene  
 13- naphthalene  
 14- hexachlorobutadiene  
 15- hexachlorocyclopentadiene  
 16- 2-chloronaphthalene  
 17- acenaphthalene  
 18- dimethyl phthalate  
 19- 2,6-dinitrotoluene  
 20- acenaphthene  
 21- 2,4-dinitrotoluene  
 22- fluorene  
 23- 4-chlorophenyl phenyl ether

24- diethyl phthalate  
 25- azobenzene  
 26- n-nitrosodiphenyl amine  
 27- 4-bromophenyl phenyl ether  
 28- hexachlorobenzene  
 29- phenanthrene  
 30- anthracene  
 31- di-n-butyl phthalate  
 32- fluoranthene  
 33- pyrene  
 34- benzidine  
 35- butyl benzyl phthalate  
 36- benzo (a) anthracene  
 37- chrysene  
 38- 3,3'-dichlorobenzidine  
 39- bis(2-ethylhexyl) phthalate  
 40- di-n-octyl phthalate  
 41- benzo (b) fluoranthene  
 42- benzo (k) fluoranthene  
 43- benzo (a) pyrene  
 44- indeno (1,2,3- $\alpha$ ) pyrene  
 45- dibenzo (a,h) anthracene  
 46- benzo (g,h,i) perylene

